



TEST CERTIFICATION

This is to certify that the I-66NL chamber serial number 32801.18.23 provided by Percival Scientific, Inc. for Percival Scientific Hong Kong Co., Ltd. has successfully tested to meet the specifications listed below:

- The chamber has met the temperature range test of 2°C to 44°C ± 0.5°C *
- The chamber has met the humidity range test of down to 30% ± 5% @ 4°C *

* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.

Testing was performed with a calibrated circular chart recorder.

Date 11/06/23