



TEST CERTIFICATION

This is to certify that the I-41VL chamber serial number 33586.05.24 provided by Percival Scientific, Inc. for Oriental Science & Technology Co. (OSTC) has successfully tested to meet the specifications listed below:

✓ The chamber has met the temperature range test of 2°C to $44^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$ (lights off) *
 4°C to $44^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$ (lights on) *

✓ The chamber has met the humidity range test of 40% to 85% \pm 10% (lights off) *

* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.

Testing was performed with a calibrated circular chart recorder.

Date 04/03/2024