



TEST CERTIFICATION

This is to certify that the CU-41L4 chamber serial number 33586.06.24 provided by Percival Scientific, Inc. for Percival Scientific Hong Kong Co. Ltd. has successfully tested to meet the specifications listed below:

✓ The chamber has met the temperature range test of 2°C to 44°C ± 0.5°C (lights off) *
10°C to 44°C ± 0.5°C (lights on) *

✓ The chamber has met the humidity range test of 40 to 85% (lights off) *

* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.

Testing was performed with a calibrated circular chart recorder.

Date 05/06/2024