



TEST CERTIFICATION

This is to certify that the E-41L2 chamber serial number 33586.09.24 provided by Percival Scientific, Inc. for Percival Scientific Hong Kong Co., Ltd. has successfully tested to meet the specifications listed below:

✓ The chamber has met the temperature range test of 2°C to 44°C ± 0.5°C (lights off) *
7°C to 44°C ± 0.5°C (lights on) *

✓ The chamber has met the humidity range test of 40 to 85% (lights off) *

* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.

Performance testing was performed at the Percival Scientific factory prior to shipment. The chamber was calibrated against an NIST traceable device. Data acquisition devices used for testing were calibrated to the same NIST standard.

Date 4/15/24