



## TEST CERTIFICATION

This is to certify that the GR-36L chamber serial number 34130.01.24 provided by Percival Scientific, Inc. for Percival Scientific Hong Kong Co., Ltd. has successfully tested to meet the specifications listed below:

✓ The chamber has met the temperature range test of 2°C to 44°C ± 0.5°C (lights off) \*  
5°C to 44°C ± 0.5°C (lights on) \*

N/A The chamber has met the humidity range test of No humidity system ordered

\* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.

Testing was performed with a calibrated circular chart recorder.

Date 7/5/2024