

TEST CERTIFICATION

This is to certify that the <u>ARC-36L2-E</u> chamber serial number <u>34275.01.24</u> provided by <u>Percival Scientific, Inc.</u> for <u>Percival Scientific Hong Kong Co., Ltd.</u> has successfully tested to meet the specifications listed below:	
The chamber has met the temperature range test of	
-	7° C to 44° C $\pm 0.5^{\circ}$ C (lights on) *
N/A The chamber has met the humidity range test of	No humidity system ordered
* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.	
Testing was performed with a calibrated circular chart recorder.	
Date08/27/2024	