

TEST CERTIFICATION

| This is to certify that the <u>ARC-36L2-E</u> chamber serial number <u>34275.03.24</u> provided by <u>Percival Scientific, Inc.</u> for <u>Percival Scientific Hong Kong Co., Ltd.</u> has successfully tested to meet the specifications listed below: |
|---|
| The chamber has met the temperature range test of 2°C to $44^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$ (lights off) * |
| 7° C to 44° C $\pm 0.5^{\circ}$ C (lights on) * |
| N/A The chamber has met the humidity range test of No humidity system ordered |
| * - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges. |
| Testing was performed with a calibrated circular chart recorder. |
| Date08/27/2024 |