

TEST CERTIFICATION

This is to certify that the <u>ARC-36L2-E</u> chamber serial number <u>34275.04.24</u> provided by <u>Percival Scientific, Inc.</u> for <u>Percival Scientific Hong Kong Co., Ltd.</u> has successfully tested to meet the specifications listed below:	
The chamber has met the temperature range test of 2° C to	
7°C to	$0.44^{\circ}\text{C} \pm 0.5^{\circ}\text{C} \text{ (lights on)} *$
N/A The chamber has met the humidity range test of No 1	humidity system ordered
* - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges.	
Testing was performed with a calibrated circular chart recorder.	
Date08/27/2024	