

TEST CERTIFICATION

| This is to certify that the <u>ARC-36L2-E</u> chamber serial number <u>34275.05.24</u> provided by <u>Percival Scientific, Inc.</u> for <u>Percival Scientific Hong Kong Co., Ltd.</u> has successfully tested to meet the specifications listed below: | |
|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|------------------------------------------------------------------------------------------------------------------------------|
| \checkmark The chamber has met the temperature range test of 2° C | to $44^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$ (lights off) * to $44^{\circ}\text{C} \pm 0.5^{\circ}\text{C}$ (lights on) * |
| <u>/ C</u> | 10 44 C ± 0.3 C (lights on) |
| N/A The chamber has met the humidity range test of No. | humidity system ordered |
| * - These temperatures and humidity levels are achievable when ambient conditions are 24°C (75°F) or below. Higher ambient temperatures may affect the operating ranges. | |
| Testing was performed with a calibrated circular chart recorder. | |
| Date08/27/2024 | |